

Certificate of Analysis

Salmonella Group B Antibody Test Kit

Product Code: CK118

Lot No: **FS9563**

Expiry Date: **19th January 2025**

	Kit Component	Lot No		Kit Component	Lot No
1	Coated Plates	CH7835	6	Sample Diluent Reagent	CH8195
2	Conjugate Reagent	CH8469	7	Wash Buffer Powder	CH6318
3	Substrate Tablets	CH8406	8	Negative Control	CH6861
4	Substrate Reagent	CH8412	9	Positive Control	CH8470
5	Stop Solution	CH8335			

Quality Control Results

Mean Negative Control

Optical Density = 0.091

Mean Positive Control

Optical Density = 0.826

Pre-diluted Reference Control
(Product Code CR100, Lot no. RF18)

Antibody titre = 693
(range: 654 – 2000)

Date of test: 19th January 2024

The above result was generated by performing the test according to the package insert.

The above information is correct and forms part of the BioChek Quality Control for Salm Gp B ELISA kit product code CK118.



Released by QC – 23rd January 2024



Jaroslaw Wnuk
Quality Control Manager