

Certificate of Analysis

Infectious Laryngotracheitis Antibody Test Kit

Product Code: CK124

Lot No: **FS9599**

Expiry Date: **2nd February 2025**

	Kit Component	Lot No		Kit Component	Lot No
1	Coated Plates	CH8409	6	Sample Diluent Reagent	CH8391
2	Conjugate Reagent	CH8543	7	Wash Buffer Powder	CH6318
3	Substrate Tablets	CH8406	8	Negative Control	CH6861
4	Substrate Reagent	CH8445	9	Positive Control	CH8544
5	Stop Solution	CH8504			

Quality Control Results

Mean Negative Control

Optical Density = 0.110

Mean Positive Control

Optical Density = 0.673

Pre-diluted Reference Control
(Product Code CR100, Lot no. RF18)

Antibody titre = 2921
(range: 2000 – 5000)

Pre-diluted Reference Control
(Product Code CR100, Lot no. RF19)

Antibody titre = 3452
(range: 3000 – 6500)

Date of test: 2nd February 2024

The above result was generated by performing the test according to the package insert.

The above information is correct and forms part of the BioChek Quality Control for ILT ELISA kit product code CK124.



Released by QC –14th February 2024



Jaroslaw Wnuk
Quality Control Manager