

Certificate of Analysis

Fowl Adenovirus Group 1 Antibody Test Kit

Product Code: CK132

Lot No: FS9605

Expiry Date: 16th February 2025

	Kit Component	Lot No		Kit Component	Lot No
1	Coated Plates	CH8434	6	Sample Diluent Reagent	CH8516
2	Conjugate Reagent	CH8582	7	Wash Buffer Powder	CH6318
3	Substrate Tablets	CH8524	8	Negative Control	CH6861
4	Substrate Reagent	CH8445	9	Positive Control	CH8583
5	Stop Solution	CH8504			

Quality Control Results

Mean Negative Control

Optical Density = 0.149

Mean Positive Control

Optical Density = 0.852

Pre-diluted Reference Control
(Product Code CR100, Lot no. RF18)

Antibody titre = 2209
(range: 1500 – 5000)

Pre-diluted Reference Control
(Product Code CR100, Lot no. RF19)

Antibody titre = 4661
(range: 3500 – 8000)

Freeze-Dried Reference Serum
(Product Code CR300, Lot no. RS08)

Antibody titre = 8425

Date of test: 16th February 2024

The above result was generated by performing the test according to the package insert.

The above information is correct and forms part of the BioChek Quality Control for FAdV ELISA kit product code CK132.



Released by QC – 23rd February 2024



Andrew Russell
Quality Control Supervisor